

Table de Conversion sonde Platine PT1000

° (C)	R (Ω)	° (C)	R (Ω)	° (C)	R (Ω)	° (C)	R (Ω)	° (C)	R (Ω)	° (C)	R (Ω)
100	1388.8	50	1194	0	1000	-50	803.1	-100	602.5	-150	397.1
99	1381.2	49	1190.1	-1	996.1	-51	799.1	-101	598.5	-151	393
98	1377.4	48	1186.2	-2	992.2	-52	795.1	-102	594.4	-152	388.8
97	1373.6	47	1182.4	-3	988.3	-53	791.1	-103	590.4	-153	384.6
96	1369.8	46	1178.5	-4	984.4	-54	787.2	-104	586.3	-154	380.4
95	1366	45	1174.7	-5	980.4	-55	783.2	-105	582.2	-155	376.3
94	1362.2	44	1170.8	-6	976.5	-56	779.2	-106	578.2	-156	372.1
93	1358.4	43	1167	-7	972.6	-57	775.2	-107	574.1	-157	367.9
92	1354.6	42	1163.1	-8	968.7	-58	771.3	-108	570	-158	363.7
91	1350.8	41	1159.3	-9	964.8	-59	767.3	-109	566	-159	359.5
90	1347	40	1155.4	-10	960.9	-60	763.3	-110	561.9	-160	355.3
89	1343.2	39	1151.5	-11	956.9	-61	759.3	-111	557.8	-161	351.1
88	1339.4	38	1147.7	-12	953	-62	755.3	-112	553.8	-162	346.9
87	1335.6	37	1143.8	-13	949.1	-63	751.3	-113	549.7	-163	342.7
86	1331.8	36	1139.9	-14	945.2	-64	747.3	-114	545.6	-164	338.5
85	1328	35	1136.1	-15	941.2	-65	743.3	-115	541.5	-165	334.3
84	1324.2	34	1132.2	-16	937.3	-66	739.3	-116	537.4	-166	330.1
83	1320.4	33	1128.3	-17	933.4	-67	735.3	-117	533.3	-167	325.9
82	1316.6	32	1124.5	-18	929.5	-68	731.3	-118	529.2	-168	321.6
81	1312.7	31	1120.6	-19	925.5	-69	727.3	-119	525.2	-169	317.4
80	1308.9	30	1116.7	-20	921.6	-70	723.3	-120	521.1	-170	313.2
79	1305.1	29	1112.8	-21	917.7	-71	719.3	-121	517	-171	309
78	1301.3	28	1109	-22	913.7	-72	715.3	-122	512.9	-172	304.7
77	1297.5	27	1105.1	-23	909.8	-73	711.3	-123	508.8	-173	300.5
76	1293.7	26	1101.2	-24	905.9	-74	707.3	-124	504.7	-174	296.3
75	1289.8	25	1097.3	-25	901.9	-75	703.3	-125	500.6	-175	292
74	1286	24	1093.5	-26	898	-76	699.3	-126	496.4	-176	287.8
73	1282.2	23	1089.6	-27	894	-77	695.3	-127	492.3	-177	283.5
72	1278.4	22	1085.7	-28	890.1	-78	691.3	-128	488.2	-178	279.3
71	1274.5	21	1081.8	-29	886.2	-79	687.3	-129	484.1	-179	275
70	1270.7	20	1077.9	-30	882.2	-80	683.3	-130	480	-180	270.8
69	1266.9	19	1074	-31	878.3	-81	679.2	-131	475.9	-181	266.5
68	1263.1	18	1070.2	-32	874.3	-82	675.2	-132	471.8	-182	262.3
67	1259.2	17	1066.3	-33	870.4	-83	671.2	-133	467.6	-183	258
66	1255.4	16	1062.4	-34	866.4	-84	667.2	-134	463.5	-184	253.7
65	1251.6	15	1058.5	-35	862.5	-85	663.1	-135	459.4	-185	249.4
64	1247.7	14	1054.6	-36	858.5	-86	659.1	-136	455.2	-186	245.2
63	1243.9	13	1050.7	-37	854.6	-87	655.1	-137	451.1	-187	240.9
62	1240.1	12	1046.8	-38	850.6	-88	651.1	-138	447	-188	236.6
61	1236.2	11	1042.9	-39	846.7	-89	647	-139	442.8	-189	232.3
60	1232.4	10	1039	-40	842	-90	643	-140	438.7	-190	228
59	1228.6	9	1035.1	-41	838.8	-91	639	-141	434.5	-191	223.7
58	1224.7	8	1031.2	-42	834.8	-92	634.9	-142	430.4	-192	219.4
57	1220.9	7	1027.3	-43	830.8	-93	630.9	-143	426.3	-193	215.1
56	1217	6	1023.4	-44	826.9	-94	626.8	-144	422.1	-194	210.8
55	1213.2	5	1019.5	-45	822.9	-95	622.8	-145	417.9	-195	206.5
54	1209.3	4	1015.6	-46	818.9	-96	618.7	-146	413.8	-196	202.2
53	1205.5	3	1011.7	-47	815	-97	614.7	-147	409.6	-197	197.9
52	1201.6	2	1007.8	-48	811	-98	610.6	-148	405.5	-198	193.6
51	1197.8	1	1003.9	-49	807	-99	606.6	-149	401.3	-199	189.3
										-200	184.9